

## The Accuracy in Measurement of Blood Pressure (AIM#BP) collaborative: Background and rationale

Item Type	Article
Authors	Padwal, R;Campbell, N.R.C;Weber, M.A;Lackland, D;Shimbo, D;Zhang, X-H;Schutte, A.E;Rakotz, M;Wozniak, G;Townsend, R;McManus, R;Asayama, K;Picone, D;Cohen, J;Brady, T;Hecht# Olsen, M;Delles, C;Alpert, B;Dart, R;DiPette, D.J;Sharman, J.E
Citation	Padwal R, Campbell NRC, Weber MA, Lackland D, Shimbo D, Zhang XH, Schutte AE, Rakotz M, Wozniak G, Townsend R, McManus R, Asayama K, Picone D, Cohen J, Brady T, Hecht-Olsen M, Delles C, Alpert B, Dart R, DiPette DJ, Sharman JE; Accuracy in Measurement of Blood Pressure Collaborative. The Accuracy in Measurement of Blood Pressure (AIM-BP) collaborative: Background and rationale. J Clin Hypertens (Greenwich). 2019 Dec;21(12):1780-1783. doi: 10.1111/jch.13735.
DOI	<a href="https://doi.org/10.1111/jch.13735">10.1111/jch.13735</a>
Journal	Journal of Clinical Hypertension
Rights	Attribution 3.0 United States
Download date	2025-01-16 12:07:54
Item License	<a href="http://creativecommons.org/licenses/by/3.0/us/">http://creativecommons.org/licenses/by/3.0/us/</a>
Link to Item	<a href="https://www.ncbi.nlm.nih.gov/pmc/articles/PMC8030396/">https://www.ncbi.nlm.nih.gov/pmc/articles/PMC8030396/</a>

**POLICY STATEMENT FROM THE  
WORLD HYPERTENSION LEAGUE****The Accuracy in Measurement of Blood Pressure (AIM-BP)  
collaborative: Background and rationale**

Raj Padwal MD<sup>1</sup>  | Norm R. C. Campbell MD<sup>2</sup>  | Michael A. Weber MD<sup>3</sup> |  
 Daniel Lackland DrPH<sup>4</sup> | Daichi Shimbo MD<sup>5</sup> | Xin-Hua Zhang MD<sup>6</sup> |  
 Aletta E. Schutte PhD<sup>7</sup> | Michael Rakotz MD<sup>8</sup>  | Gregory Wozniak PhD<sup>8</sup> |  
 Raymond Townsend MD<sup>9</sup> | Richard McManus MD<sup>10</sup> | Kei Asayama MD<sup>11,12,13</sup> |  
 Dean Picone PhD<sup>14</sup> | Jordy Cohen MD<sup>15</sup> | Tammy Brady MD<sup>16</sup> | Michael Hecht-Olsen MD<sup>17</sup> |  
 Christian Delles MD<sup>18</sup> | Bruce Alpert MD<sup>19</sup> | Richard Dart MD<sup>20</sup> |  
 Donald J. DiPette PhD<sup>21</sup>  | James E. Sharman PhD<sup>14</sup> | for the Accuracy in Measurement  
 of Blood Pressure Collaborative

<sup>1</sup>Department of Medicine, University of Alberta, Edmonton, AB, Canada

<sup>2</sup>Department of Medicine, Physiology and Pharmacology, and Community Health Sciences, O'Brien Institute for Public Health and Libin Cardiovascular Institute of Alberta, University of Calgary, Calgary, AB, Canada

<sup>3</sup>Division of Cardiovascular Medicine, SUNY Downstate Medical Center, Brooklyn, NY, USA

<sup>4</sup>Division of Translational Neurosciences and Population Studies, Department of Neurology, Medical University of South Carolina, Charleston, SC, USA

<sup>5</sup>Columbia Hypertension Center, Columbia University Medical Center, New York, NY, USA

<sup>6</sup>World Hypertension League, Beijing, China

<sup>7</sup>MRC Unit for Hypertension and Cardiovascular Disease, Hypertension in Africa Research Team (HART), North-West University, Potchefstroom, South Africa

<sup>8</sup>American Medical Association, Chicago, IL, USA

<sup>9</sup>Perelman School of Medicine, University of Pennsylvania, Philadelphia, PA, USA

<sup>10</sup>Nuffield Department of Primary Care Health Sciences, Radcliffe Observatory Quarter, University of Oxford, Oxford, UK

<sup>11</sup>Department of Hygiene and Public Health, Teikyo University School of Medicine, Tokyo, Japan

<sup>12</sup>Tohoku Institute for Management of Blood Pressure, Sendai, Japan

<sup>13</sup>Research Unit Hypertension and Cardiovascular Epidemiology, Department of Cardiovascular Sciences, University of Leuven, KU Leuven, Leuven, Belgium

<sup>14</sup>Menzies Institute for Medical Research, University of Tasmania, Australia

<sup>15</sup>Renal-Electrolyte and Hypertension Division, Perelman School of Medicine, University of Pennsylvania, Philadelphia, PA, USA

<sup>16</sup>Division of Pediatric Nephrology, Johns Hopkins University School of Medicine, Baltimore, MD, USA

<sup>17</sup>Department of Internal Medicine, Centre for Individualized Medicine in Arterial Disease, Holbaek Hospital, University of Southern, Odense, Denmark

<sup>18</sup>Institute of Cardiovascular and Medical Sciences, University of Glasgow, Glasgow, UK

<sup>19</sup>Division of Cardiology, University of Tennessee Health Science Center, Memphis, TN, USA

<sup>20</sup>Center for Precision Medicine and Research, Marshfield Clinic Research Institute, Marshfield, WI, USA

<sup>21</sup>Health Sciences Distinguished Professor, University of South Carolina School of Medicine, Columbia, SC, USA

**Correspondence** Raj Padwal, Clinical Epidemiology, Clinical Pharmacology and General Internal Medicine, Professor of Medicine and Director, Hypertension Clinic, University of Alberta, 5-134A Clinical Sciences Building, 11350 - 83rd Ave, Edmonton, AB T6G 2G3, Canada.

Email: rpadwal@ualberta.ca

**Funding information**

This project is unfunded. Raj Padwal is Co-Founder and CEO of mmHg Inc, a University of Alberta based startup creating software and hardware solutions to improve blood pressure measurement. Richard J McManus has received BP monitors for research purposes from Omron. Norm Campbell was a paid consultant to the Novartis Foundation (2016-2017) to support their program to improve hypertension control in low- to middle-income countries which includes travel support for site visits and a contract to develop a survey; has provided paid consultative advice on accurate blood pressure assessment to Midway Corporation (2017); and is an unpaid member of World Action on Salt and Health (WASH). Bruce Alpert is a consultant for development of BP technology for Welch Allyn/

Hill Rom, Midmark, and Aktia. Donald DiPette serves as a consultant to the Centers for Disease Control (CDC), World Health Organization (WHO), and the Pan American Health Organization (PAHO). Michael Hecht-Olsen received a part-time research grant from Novo Nordic Foundation (2013-2018). Michael Rakotz and Gregory Wozniak are employees of the American Medical Association. Aletta Schutte has received speaker honoraria from Novartis, Omron, Takeda, and Servier and paid consultative advice to Abbot Pharmaceuticals. She is funded by the South African Medical Research Council and National Research Foundation (SARChI). James Sharman has no personal disclosures; however, his university has received equipment and research funding from manufacturers of BP devices including AtCor Medical, IEM, and Pulsecor (Uscom). Kei Asayama is a consultant for Omron Healthcare. Richard McManus is Chair of the British and Irish Hypertension Society BP measurement working party and has received BP monitors for research use from Omron, and his institution has received research grants from The National Institute for Health Research (NIHR) and Stroke Association/British Heart Foundation for work on self-monitoring of hypertension.

Blood pressure (BP) measurement, a technique first described over a century ago, is an essential component of clinical care and critical for the detection and management of hypertension.<sup>1</sup> Accordingly, the ramifications of inaccurate BP measurement, which is a persistent and pervasive problem worldwide, are profound.<sup>2</sup> Assuming a global prevalence of hypertension of 1.4 billion,<sup>3</sup> a 5-mmHg error in BP measurement has been estimated to result in the incorrect classification of hypertension status in at least 84 million individuals worldwide.<sup>4</sup> In addition, incorrect classification has important ramifications for individual patients, whether it leads to misdiagnosis and inappropriate prescribing of antihypertensive drugs, or to lack of recognition of a clinical condition that can cause devastating cardiovascular consequences. The continued rise in the global prevalence of hypertension has made work related to optimizing BP measurement even more critical, and notwithstanding similar initiatives that have been conducted or are ongoing, additional efforts to improve BP measurement on a global scale are clearly needed.<sup>5</sup>

The Accuracy in Measurement of BP (AIM-BP) Collaborative, which has been formed to advance and amplify past and current work, is focused on advocating for and implementing optimal BP measurement practices globally. This collaborative is intended to be complimentary, and not competitive to, existing efforts in the field of BP measurement. The Lancet Commission on Hypertension Group and the World Hypertension League are lead organizations. AIM-BP is comprised of academics, clinicians, professional societies, non-governmental organizations, and additional stakeholders that have expertise and interest in the field of BP measurement. An advisory group of members from industry has also been created because these stakeholders can make important contributions, but their input remains independent of priority setting. Three major initial objectives have been identified: 1. publish a series of articles in the *Journal of Clinical Hypertension* detailing developments and challenges in the field; 2. host a webpage on the World Hypertension League website that summarizes critical information and resources for clinicians and patients; and 3. advocate for policies that are supportive of accurate BP assessment (eg, certification of training in BP assessment, accreditation of health facilities, mandating use of validated BP devices, etc). It is expected that this work will catalyze further initiatives and collaborative efforts moving forward.

Many unresolved issues and challenges exist in the field (Table 1). Lack of performance of standardized measurement in clinical settings continues to be a major contributor to inaccurate BP measurement.<sup>4</sup> Use of validated, automated devices and protocolized technique is recommended to simplify the measurement process,

ensuring closer adherence to recommended practice. Training of observers (healthcare providers measuring BP) is critical but needs to be done efficiently and with little or no added cost. An example is to use short, simple, web-based, whiteboard animation educational videos to summarize the essentials of BP measurement. These can be quickly viewed and reviewed as needed. A series of videos on this topic has been produced at Johns Hopkins University, and these are easily and freely accessible (<https://www.youtube.com/watch?v=T9J3RE4Eins>). Dissemination of these resources and translation into multiple languages are needed to reach the global audience. Similar videos could be made to provide guidance on related topics in BP measurement—such as describing how best to implement a BP measurement station in a busy clinic to maximize accuracy and clinical efficiency.

Another major issue of interest is use of automated BP measurement, which is the primary method recommended for simplifying the BP measurement process. Clinical validation of these oscillometric BP measurement devices to confirm accuracy relative to blinded, two-observer auscultation should be performed before such devices are released into the market. Unfortunately, this is not mandated by global regulatory agencies, and consequently, many devices available worldwide are not validated.<sup>6</sup> The widespread availability of unvalidated oscillometric devices is a major barrier to accurate BP measurement. Additional important issues that require clarification (Table 1) include achieving consensus on which clinical validation protocols should be used, determining how best to make validated device listings available to clinicians and patients in a user-friendly format that is locally relevant, and empowering local contract research organizations to perform validations at a cost that does not deter manufacturers from performing such studies.

With increasing emphasis being placed on out-of-office BP measurement as the preferred means of diagnosing and following up patients with hypertension, there is a need to ensure that recommended procedures for performing home BP and 24-hour ambulatory BP monitoring are followed.<sup>7-9</sup> Recently published data have confirmed that patients do not often follow proper home BP measurement and reporting protocols.<sup>10</sup> While a substantial amount of effort has been invested in outlining procedures for out-of-office BP measurement, relatively little time has been spent on implementing mechanisms that enable these procedures to be done correctly and this is expected to be an additional focus of AIM-BP.

The need for an easy to access, free central repository for the general public, patients, and physicians that summarizes reliable and essential information and displays links to useful resources is clear. Providing a venue to promote regional, national,

**TABLE 1** AIM-BP collaborative priority areas in BP measurement

Category	Item
Automated (oscillometric) devices	<ol style="list-style-type: none"> <li>1. Clinical validation               <ol style="list-style-type: none"> <li>a. Improving validation standards and device accuracy assessment</li> <li>b. Adopting and implementing new standards and retiring old protocols</li> <li>c. Comparing/contrasting existing and forthcoming regional and global validated device listings</li> <li>d. Identifying and removing barriers to performing validation, particularly in low- to middle-income settings (eg, cost, insufficient training)</li> <li>e. Identifying reputable centers in different countries that perform validation studies.</li> <li>f. Facilitating access to validation information for end-users (clinicians, patients, general public, researchers)</li> <li>g. Assessing accuracy in special populations (eg, children, pregnancy, arrhythmia, vascular disease)</li> </ol> </li> <li>2. Strengthening regulatory requirements for approval</li> <li>3. Identifying best practices for static calibration</li> </ol>
Self-measurement of BP	<ol style="list-style-type: none"> <li>1. Ensuring optimal technique (including proper cuffing)</li> <li>2. Improving device accuracy in individual patients               <ol style="list-style-type: none"> <li>a. Need for individualized algorithms</li> <li>b. Patient-specific calibration</li> <li>c. Emerging modalities (e.g., “cuffless” devices and continuous measurement)</li> </ol> </li> <li>3. Optimizing recording, summarizing, reporting, and interpreting</li> <li>4. Removing barriers to access</li> </ol>
Measurement by observers (clinicians)	<ol style="list-style-type: none"> <li>1. Ensuring recommended technique (in-office, ABPM, in-pharmacy).</li> <li>2. Providing training, certification, and ongoing maintenance of competence in a format that is free, accessible, and efficient.</li> <li>3. Using team-based care and telemonitoring.</li> <li>4. Transitioning care from childhood to adulthood</li> </ol>
Global initiatives and sharing of best practices	<ol style="list-style-type: none"> <li>1. Promote regional, national, and international BP measurement accuracy improvement initiatives</li> <li>2. Advocate for governmental regulation of devices and policies to require certification of health care professionals</li> <li>3. Advocate for changes to health care facility accreditation that facilitate accurate BP assessments</li> <li>4. Detail and discuss reimbursement-related initiatives</li> <li>5. Discuss regional and country-specific challenges in BP measurement.</li> </ol>

and international initiatives aimed at improving BP measurement is also a necessary element for a successful program. In this regard, the participation of the Lancet Commission on Hypertension Group, Journal of Clinical Hypertension, and World Hypertension League in the AIM-BP collaborative is of critical importance because these entities are committed to the open access of information and dedicated to improving hypertension on a global level. It is important that information be presented in an actionable and user-friendly format.

A foundational element of AIM-BP is that it is an inclusive and collaborative initiative. As such, membership is open, and we invite individuals or entities with an interest in promoting evidence-based, accurate BP measurement to join. Although much needs to be done, we anticipate that our combined efforts will result in improved BP measurement and, ultimately, better diagnosis and assessment of BP-related conditions on a global scale.

## AIM-BP ORGANIZATIONAL MEMBERS

World Hypertension League (lead), Lancet Commission on Hypertension Group (lead), American Medical Association, Hypertension Canada, British and Irish Society of Hypertension, Japanese Society of Hypertension, and Danish Hypertension Society.

## AIM-BP INDUSTRY ADVISORY MEMBERS

David Quinn (Hillrom), Josh Sarkis (PharmaSmart), and Jim Li (Omron).

### ORCID

Raj Padwal  <https://orcid.org/0000-0003-3541-2817>

Norm R. C. Campbell  <https://orcid.org/0000-0002-1093-4742>

Michael Rakotz  <https://orcid.org/0000-0003-4386-4864>

Donald J. DiPette  <https://orcid.org/0000-0002-5762-9104>

### REFERENCES

1. Booth J. A short history of blood pressure measurement. *Proc R Soc Med.* 1977;70:793-799.
2. Olsen MH, Angell SY, Asma S, et al. A call to action and a lifecourse strategy to address the global burden of raised blood pressure on current and future generations: the lancet commission on hypertension. *Lancet.* 2016;388:2665-2712.
3. Mills KT, Bundy JD, Kelly TN, et al. Global disparities of hypertension prevalence and control: a systematic analysis of population-based studies from 90 countries. *Circulation.* 2016;134:441-450.
4. Padwal R, Campbell NRC, Schutte AE, et al. Optimizing observer performance of clinic blood pressure measurement: a position

- statement from the Lancet Commission on Hypertension Group. *J Hypertens*. 2019;37(9):1737-1745.
5. Kallioinen N, Hill A, Horswill MS, Ward HE, Watson MO. Sources of inaccuracy in the measurement of adult patients' resting blood pressure in clinical settings. *J Hypertens*. 2017;35:421-441.
  6. Alpert BS. Can 'FDA-cleared' blood pressure devices be trusted? A call to action. *Blood Press Monit*. 2017;22:179-181.
  7. Whelton PK, Carey RM, Aronow WS, et al. 2017 ACC/AHA/AAPA/ABC/ACPM/AGS/APhA/ASH/ASPC/NMA/PCNA Guideline for the prevention, detection, evaluation, and management of high blood pressure in adults: a report of the American college of cardiology/American heart association task force on clinical practice guidelines. *Hypertension*. 2018;71:e13-e115.
  8. Nerenberg KA, Zarnke KB, Leung AA, et al. Hypertension Canada's 2018 guidelines for diagnosis, risk assessment, prevention, and treatment of hypertension in adults and children. *Can J Cardiol*. 2018;34:506-525.
  9. National Institute for Health and Care Excellence. *Hypertension in Adults: Diagnosis and Management*. Clinical Guideline CG127. London, UK: National Institute for Health and Care Excellence;2011. (Updated 2016). Accessed on July 12, 2019. <https://www.nice.org.uk/guidance/cg127>
  10. Milot JP, Birnbaum L, Larochelle P, et al. Unreliability of home blood pressure measurement and the effect of a patient-oriented intervention. *Can J Cardiol*. 2015;31:658-663.

**How to cite this article:** Padwal R, Campbell NRC, Weber MA, et al; for the Accuracy in Measurement of Blood Pressure Collaborative. The Accuracy in Measurement of Blood Pressure (AIM-BP) collaborative: Background and rationale. *J Clin Hypertens*. 2019;00:1-4. <https://doi.org/10.1111/jch.13735>